



**City of Rocky Mount
Addendum 1**

Refer ALL Inquiries regarding this RFQ to:
Tracy Johnson, Purchasing Clerk
tracy.johnson@rockymountnc.gov

RFP #320-280722TJ – Segment Flow Analyzer

RFQ Due Date: 9/1/2022

Contract Type: Goods

1. Question: Does the system need to be able to analyze both the Total Phosphate and Total Nitrogen Simultaneously?

Answer: It is not a requirement. As long as all the methods listed in the RFP are covered. Generally we do not run the machine in dual channel mode.

2. Question: Does the modules for Total Nitrogen and Total Phosphate require inline digestion? If not, does the City want this quoted as optional add ons?

Answer: Currently we digest separately and then analyze. It is unlikely that we would switch to inline digestion but quoting it as an optional addon is fine.

3. Question: Should we provide a quote for a TKN manifold that can also run ammonia via gas diffusion as an optional add on as well?

Answer: Quoted as an addon is fine.

4. Question: If optional add on products are offered is it desirable for Skalar to allowing pricing for 365 days?

Answer: If there is a significant increase in price to offer 365 day pricing it would most likely be better to offer a standard quote period. Chances that we would come back to buy addons later are low.

Overall, it would be best to include an itemized quote so the City may make a selection if needed.

Vendor: _____